Data sheet

100 MHz Analog Oscilloscope

Model 2190B

- Dual time base oscilloscope (2 channel)
- 5 mV/division sensitivity
- Sweeps to 5 ns/division
- 23 calibrated ranges, main time base
- Signal delay line
- 12 kV accelerating voltage
- Channel 2 output



Specifications	
	2190B
VERTICAL AMPLIFIERS (CI	H I and CH 2)
Sensitivity	5 mV/div to 5 V/div. 1 mV/div to 1V/div (at X5 MAG
Attenuator	10 calibrated steps in 1-2-5 sequence.
	Vernier control provides fully adjustable sensitivity
	between steps, adjustment range 1/1 to 1/3
Accuracy	±3% (±5% at X5 MAG)
Input Impedance	I MΩ +2%
Input Capacitance	25 pF ±10 pF
Frequency Response	DC: DC to 100 MHz (-3 dB)
X5 MAG	DC to 25 MHz (-3 dB)
AC	10 Hz to 100 MHz (-3 dB)
Rise Time	3.5 ns (Overshoot ≤5%)
Signal Delay Time	Variable
Square Wave Characteristics	Overshoot less than 5%, 10 mV/div range
	Other ranges within 5% additional
Maximum Input Voltage	400 V (DC + AC peak)
VERTICAL AMPLIFIERS	
Operating Modes	CH 1, CH 2, Dual, Add
Delay Time Between Channels	Within I ns between CH I and CH 2
Crosstalk	30:1 at 100 kHz
SWEEP SYSTEM	
Operating Modes	
A	A sweep
В	Delayed B sweep
B TRIGGERED	B sweep triggered after delay
A Time Base	
Sweep Mode	Main, Mix, Delay, XY
Sweep Time:	5 s to 20 ns/div., 23 steps in 1-2-5 sequence
	with variable control
Accuracy	± 3%
Hold Off Time	Continuously variable. Adjustment range from
	normal to 5 times normal
B Time Base	
Delay Method	Continuous delay. Triggered delay
Sweep Time	20 ns. to 0.5 s/div., 23 steps in 1-2-5 sequence
Accuracy	± 3%
Delay Time	Start point: 0.5 div to + 0.3 div.
	End point: 10 div + 1 div
Delay Jitter	Within 1/10,000 of full scale sweep time
TRIGGERING	
A Trigger	
Source	CH 1, CH 2, LINE, EXT, ALT
Sensitivity	30 Hz to 110 MHz
	1.5 div (internal), ≥0.5 p-p (external)
TV-V	20 Hz - 30 kHz
	1.0 div (internal), ≥0.5 p-p (external)
TV-H	3 kHz - 100 kHz
	1.0 div (internal), ≥0.5 p-p (external)
Slope	+ or -
B Trigger	The A trigger is also the B trigger

Maximum Input Voltage	300 V (DC + AC peak)
Maximum input voltage	300 V (DC 1 /1C pcak)
IODIZONITAL AMBUIELE	n.
HORIZONTAL AMPLIFIE	
X-Y Mode	X Axis = CH I. Y Axis = CH 2
Sensitivity	5 mV/div to 5 V/div, CH I and CH 2
Accuracy	±3% calibrated position, ±6% using x10 MAG
Frequency Response	DC to 2 MHz (-3 dB)
Freq. Response	20 Hz to 100 MHz, -3 db
CH2 (Y) OUTPUT Output Voltage	Approx. 100 mV/div open circuit
	Approx. 50 mV/div into 50 Ω
	*
Output Impedance	approx. 50 Ω
CDT	
CRT	
Type	Rectangular with integral graticule
	$8 \times 10 \text{ div } (1 \text{ div} = 1 \text{ cm})$
Display Area	8 x 10 div (1 div — 1 cili)
Display Area Accelerating Voltage	12 kV
Accelerating Voltage	12 kV

Other Specifications	
Z Axis	Sensitivity: 3 V or greater, TTL level.
(Intensity Modulation)	Intensity increasing with more positive levels
Input Impedance	50 kΩ
Usable Freq. Range	DC to 5 MHz
Maximum Input Voltage	30 V (DC + AC peak)
CAL/Probe Compensation	
Waveform	Positive going squareware
Output Voltage	2 V p-p ±3%
Frequency	Approx. I kHz
Power Requirements	100/120/220/240/ VAC ±10%, 50/60 Hz,
	approximately 55 W
Dimensions (HxWxD)	12.76 x 15.68 x 5.2" (324 x 398 x 132 mm)
Weight	18.7 lbs (8.5 kg)
ENVIRONMENT	
Within Specified Accuracy	50° to 95°F (10° to 35°C), 10-80% RH
Full Operation	32° to 122°F (0° to +50°C), 10-80% RH
Storage	-22° to 158°F (-30° to +70°C), 10-90% RH
	Three Year Warranty

Accessories

Supplied: Instruction Manual, Two PR 37A x1/x10/Ref. Probes or equivalent,

AC Power Cord, Spare Fuse

Optional: PR 32A Demodulator Probe, PR 37AG x1/x10/REF. Probe,

PR 100A x100 Probe, PR-55 High Voltage x1000 Probe,

LC 210A Carrying Case

