

CAT III 1kV Fused Test Leads

P/N 131370/OR

"Fused" Test Leads Provide Added Protection for User & Instrument!



Specifications

- Compatible with all Industry Standard 4mm Instrument Sockets
- Double Insulated, Bi-Color Silicone Cable boasts 250+ copper strands for maximum flexibility and overall performance
- Fused leads protect user from shock -even if the instrument is fused!
- Fuse is easily serviced without the use of tools



Complies with/Conforme avec
UL61010-1
CSA C22.2 No. 61010-1
With/Avec EN61010-2-31

Rated 1000V CAT III
Calibré 1000V CAT III

Fuse Rating: 500mA/1000V/50kA
Calibre du fusible : 500mA/1000V/50kA



Date
19/11/01

Drn.By
Steve S

Title
OEM FUSE PROBE TO RA LEAD KIT
124013 FIXED TIP FRONT

ISSUE No.

1
19/11/01
PROD. ISSUE

2
05/12/02
Contact Change

Part No.

131370/OR

This drawing is not to be disclosed to a third party without the authority from the company.©


Scale
N.T.S.

IEC 1010 COMPLIANT RATING:
1000 VOLTS CAT III 0.5 AMPS MAX. POL. DEG. 2
0.5 AMPS CONT.

PROJECT No. -

1No. RED & BLACK 123872 LEADS

125076 OEM RA PLUG



ASSEMBLY JIG NUMBERS

PACKAGING
Poly Bag: 111500

PRINTING AS ABOVE

TOLERANCES

CABLE LENGTHS +/- 5%

DO NOT SCALE.
ALL DIMENSIONS IN mm.

Telephone: +44 (0)1992 450 330
Fax No: +44 (0)1992 450 457

Unit 42 Hoddesdon Ind.
Centre, Pindar Road
Hoddesdon, Herts.
England, EN11 0FF

DECLARATION OF CONFORMITY

SILVERTRONIC Ltd. confirm that the products listed below have been Designed, Manufactured and Tested to, and therefore are fully compliant to;
IEC1010-1 1990 inc. Amendment 2-1995 & IEC1010-2-031 First Edition-1993
and thus meet *THE LOW VOLTAGE DIRECTIVE (LVD) 73/23/EEC and amending directive 93/68/EEC*

PART No:

131370/OR

PRODUCT TYPE / DESCRIPTION:

OEM Fused Probe To OEM OR Plug

IEC 1010 RATING:

1000V CAT III (3) 600V CAT IV (4)

MAXIMUM RATED CURRENT:

0.5A

POLLUTION DEGREE:

2

INCLUDES THE FOLLOWING ASSEMBLIES / COMPONENTS SUBECT TO INDIVIDUAL TEST.

DESCRIPTION:

OEM Right Angle Plug

Fused Test Lead Assembly

PART No:

125076

123872

Approved Signatories: J.DEVEREUX (✓) A.WILLIAMS ()
Signed on behalf of **SILVERTRONIC Ltd.**

DATE: 22 March 2007



All Testing carried out by Third Party Testhouse
CRANAGE EMC TESTING Ltd.
(LVD TESTING DEPARTMENT)