

WBHC-23912-000


SHEET 2.1

DESCRIPTION

THE WBHC-23912-000 IS A MAGNETIC BALANCED ARMATURE RECEIVER INTENDED FOR USE IN ITC AND CIC HEARING INSTRUMENTS. THE WBHC FAMILY OFFERS AN EXTENDED HIGH FREQUENCY RESPONSE IN THE SAME SIZE PACKAGE AS THE HC FAMILY. ALL WBHC UNITS HAVE SHOCK PROTECTION. THIS MODEL HAS MEDIUM IMPEDANCE AND IS UNDAMPED.

NOTE: SPECIFICATIONS FOLLOWED BY AN ASTERISK (*) ARE 100% TESTED.

CONSTANT VOLTAGE DRIVE RESPONSE



ACOUSTICAL

NOTIVITY

DEVICE WILL PRODUCE THE SPL LISTED BELOW WITH THE TEST
CONDITIONS DESCRIBED IN TABLE 3. NOMINAL SENSITIVITY

AT I kHz IS dB RELATIVE TO 20 μPα. ALL OTHER VALUES IN

dB RELATIVE TO THE SENSITIVITY AT I kHz.

FREQUENCY (Hz)	MINIMUM	NOMINAL	MAXIMUM
100	-4.5	-1.5	+1.5
250	-3.0	-0.5	+2.0
500	-2.0	0.0	+2.0
1000	-2.0	95.0	+2.0
3000-3800 PEAK	+7.0	+10.0	+13.0
4100-5100 VALLEY	0.0		
4850-5850 PEAK	+3.0	+6.0	+9.0

TABLE I.

TOTAL HARMONIC DISTORTION*

DEVICE WILL NOT EXCEED TOTAL HARMONIC DISTORTION LEVELS LISTED BELOW.

FREQUENCY (Hz)	DRIVE (V RMS)	DC BIAS (MA)	LIMIT (%)
1100	0.399 V	0	5
1650	0.399 V	0	5
500	2.12 V	0	8

TABLE 2.

TEST CONDITIONS

NOMINAL SOURCE VOLTAGE	0.399 Vrms, O Vdc BIAS
SOURCE IMPEDANCE	< Ι Ω
TUBING	10 mm (.394) LONG, mm (.039) D.
COUPLER CAVITY	2 CC SIMULATED ANSI S3.7 TYPE HA-3. (IEC 126)

TABLE 3.

POSITIVE SIGNAL APPLIED TO TERMINAL 2 WILL PRODUCE A DECREASE IN SOUND PRESSURE AT THE SOUND OUTLET.

ELECTRICAL

DC RESISTANCE	275 Ω ±10%	*
IMPEDANCE @ 500 Hz	291 Ω ±15%	*
IMPEDANCE @ I kHz	327 Ω ±20%	*
INDUCTANCE @ 500Hz	9 mH ±15%	
CAPACITANCE @ 10 MHz	4 pF ±20%	

TABLE 4.


ISOLATION: THE CASE WILL BE ELECTRICALLY ISOLATED FROM THE COIL CIRCUIT \bullet

MAGNETIC RADIATION
WORST CASE: FIELD WILL BE LESS THAN LEVEL STATED BELOW
AT AMPLIFIER CLIPPING (.920 V). 124 dB re IµA/m DISTANCE OF 6.3 mm FROM CENTER OF RECEIVER ANGLE OF 45 DEGREES FROM TUBE

MECHANICAL

PORT LOCATION: 12C SOLDER TYPE: SAC305

OPERATING: SENSITIVITY WILL NOT VARY MORE THAN
+1/-3 dB FROM -17°C TO 63°C
STORAGE: -40°C TO 63°C



Revision C.O. # Implementation Date RELEASE LEVEL REVISION C10104863 10-16-06 Released

KNOWLES ELECTRONICS ITASCA, ILLINOIS U.S.A.

A	C10104036	3-31-06			
			INSPECTION ACCEPTANCE/REJECTION	DR. BY	DATE
CRITERIA, CORRELATION OF TEST EQUIPMENT WITH KNOWLES IS ALSO REQUIRED FOR ELIMINATION OF EQUIPMENT AND TEST METHOD VARIATION			AB	3-31-0	
				CK. BY	DATE
TITLE:	RF	CEIVER	WBHC-23912-000	CRG	3-30-06
	IV E	CLIVEN	#B110 20012 000	APP. BY	DATE
	PERFORMAN	NCE SPECIFICATION	SHT 2.1	ммм	3-31-06