



**NORTH AMERICAN
POWER CORDS**

Hospital Grade Power Cords





QUALTEK ELECTRONICS

Hospital Grade - Unshielded with North American Color Code

**RoHS
Compliant**



Part Number	Figure	Length	Cordage	# Cond. AWG	Amp Rating	Plug No.	Conn No.	PDF Drawings
232002-06	I	8' 0"	SJT+	3 x 18	10	5-15P*	--	
232003-06	I	10' 0"	SJT+	3 x 18	10	5-15P*	--	
232010-06	I	8' 0"	SJT+	3 x 18	10	5-15P**	--	
232011-06	I	10' 0"	SJT+	3 x 18	10	5-15P**	--	
233002-06	J	8' 0"	SJT+	3 x 18	10	5-15P*	IEC 60320 C13 (Gray)	
233003-06	J	10' 0"	SJT+	3 x 18	10	5-15P*	IEC 60320 C13 (Gray)	

233008-06	J	8' 0"	SJT+	3 x 18	10	5-15P**	IEC 60320 C13 (Gray)	
233009-06	J	10' 0"	SJT+	3 x 18	10	5-15P**	IEC 60320 C13 (Gray)	
242003-06	K	10' 0"	SJT+	3 x 18	10	5-15P*	--	
243003-06	L	10' 0"	SJT+	3 x 18	10	5-15P*	IEC 60320C13 (Gray)	

Note: -01 = Black, -06 = Gray


* Hospital Grade Clear Plug

** Hospital Grade Gray Plug

+ North American Color Code (Black/White/Green)

All Cords have 60°C and 125VAC unless marked as above

REVISION			
REV.	DESCRIPTION	DATE	APPROVED
B	RoHS Compliancy	06-01-06	D.P.



NOTES:

1. CORDAGE: SJT, 18AWG / 3C, HOSPITAL GRADE - UNSHIELDED, NA COLOR CODE, TEMP. RATING 60°C, RATING: 125V 10A, JACKET COLOR: GRAY
2. PLUG: NEMA 5-15P HOSPITAL GRADE (GRAY)
3. OTHER END: IEC 60320 C13 (GRAY)
4. APPROVALS: UL, CSA
5. ROHS COMPLIANT
6. CAUTION LABEL

CONTROLLED

Dim range (mm)	Tolerance +/- (mm)	Qualtek Electronics Corp. PPC DIVISION		
0-1	0.15	Part Number:	233008-06	RoHS Compliant
1-6	0.20	Description:	HOSPITAL GRADE POWER CORD	
6-18	0.30		UNIT: in	REV. B
18-50	0.50	Drawn / Date	Checked / Date	Approved / Date
50-120	0.70	B. ZHANG 10-08-01	10-08-01	JJH / 10-08-01
120-250	0.90			
250-500	1.20			