

Telephone: +44 (0)20 8805 3424  
Fax No: +44 (0)20 8805 3054


Leaside Industrial Estate  
Stockingswater Lane  
Enfield, Middx. England  
EN3 7PH

## DECLARATION OF CONFORMITY

SILVERTRONIC Ltd. confirm that the products listed below have been Designed, Manufactured and Tested to, and therefore are fully compliant to;  
*IEC1010-1 1990 inc. Amendment 2-1995 & IEC1010-2-031 First Edition-1993*  
and thus meet *THE LOW VOLTAGE DIRECTIVE (LVD) 73/23/EEC and amending directive 93/68/EEC*

<b>PART No:</b> <i>129303/OS</i>	
<b>PRODUCT TYPE / DESCRIPTION:</b> <i>OEM Fixed Lead Prod to Shutter Plug Test Lead Set</i>	
<b>IEC 1010 RATING:</b> <i>1000V CAT III (3)</i>	
<b>MAXIMUM RATED CURRENT:</b> <i>20A</i>	
<b>POLLUTION DEGREE:</b> <i>2</i>	
<b>INCLUDES THE FOLLOWING ASSEMBLIES / COMPONENTS SUBJECT TO INDIVIDUAL TEST.</b>	
<b>DESCRIPTION:</b>	<b>PART No:</b>
OEM Fixed Lead Prod	125074
Shutter Plug	125106
1mm sq D/S Cable	119054

Approved Signatories: J.DEVEREUX ( ) A.WILLIAMS (✓) V.NEGLIA ( )  
Signed on behalf of SILVERTRONIC Ltd.

  
\_\_\_\_\_

DATE: 25th February 2004



All Testing carried out by Third Party Testhouse  
CRANAGE EMC TESTING Ltd.  
( LVD TESTING DEPARTMENT )

**SILVERTRONIC**  
LTD.

Date  
04/03/04

Drn.By  
Vince

Title  
OEM FLP TO OS PLUG LEAD SET

ISSUE No.

1

04/03/04  
PROD. ISSUE

Part No.


**129303/OS**

This drawing is not to be disclosed to a third party without the authority from the company.©


Scale  
FULL SIZE

IEC 1010 COMPLIANT RATING:  
1000VOLTS CAT III 20AMPS POL DEG. 2

PROJECT No.



1No. RED & BLACK 123303 OEM FLP-OS LEAD



119054 CUT 1.250 Metre  
or 102523 cut length

ASSEMBLY JIG NUMBERS  
125074= J164 or 167 + 166

PACKAGING  
POLY BAG: 111500

NO PRINTING

TOLERANCES CABLE LENGTHS +/- 5%

DO NOT SCALE.  
ALL DIMENSIONS IN mm.