

SILVERTRONIC

LEADS
...the way forward!

Leaside Industrial Estate,
Stockingswater Lane,
Enfield, Middx. EN3 7PH
Telephone: +44 (0)20 8805 3424
Facsimile: +44 (0)20 8805 3054

DECLARATION OF CONFORMITY

SILVERTRONIC Ltd. confirm that the products listed below have been Designed, Manufactured and Tested to, and therefore are fully compliant to;
IEC1010-1 1990 inc. Amendment 2-1995 & IEC1010-2-031 First Edition-1993
and thus meet **THE LOW VOLTAGE DIRECTIVE (LVD) 73/23/EEC and amending directive 93/68/EEC**
GS38 Compliant

PRODUCT TYPE / DESCRIPTION				Pair If In Line Fused Links	
PART No.				134309	
IEC1010 RATING	VOLTAGE	600	Volts	INSULATION CATEGORY	III (3)

MAXIMUM RATED CURRENT	*10 as fuse	Amps
POLLUTION DEGREE	2	
INSULATION CATEGORY	DOUBLE/REINFORCED	

INCLUDES FOLLOWING ASSEMBLIES / COMPONENTS SUBJECT TO INDIVIDUAL TEST.	DESCRIPTION	PART No.
	In Line Fuse Link	121326
* Current: As fuse fitted		



All Testing carried out by Third Party Testhouse
CRANAGE EMC TESTING Ltd.
(LVD TESTING DEPARTMENT)



www.silvertronic.com
e-mail: sales@silvertronic.com

Registration No. 1993292

North American Office
4405 Sharon Copley Road,
Medina, Ohio, 44256.
Toll Free: (866) 556 3308
Facsimile: (330) 722 2281

SILVERTRONIC LTD

Title: In Line 10Amp Fuse Links

Rev	Details	Date	Drawn
4			
3			
2			
1	First issue	25/11/02	Rich
-	Pre-Production issue	25/11/02	Rich

Part No:
134309




This drawing is not to be disclosed to a third party without the written authority from the company ©

Material: ###

Project No: #####

Supplied as a Pair 1 Red and 1 Black
Individual Fuse Link Part No: 121326

Fuse Rating: 10A/600V/50KA



Assembly Jig Numbers

#####=J###

#####=J###

#####=J###

#####=J###

Packaging: Poly Bag: 111513

Tolerances:
Cable Lengths +/- ##%

Do Not Scale All Dimensions In MM