

Leaside Industrial Estate, Stockingswater Lane, Enfield, Middx. EN3 7PH Telephone: +44 (0)20 8805 3424 Facsimile: +44 (0)20 8805 3054

DECLARATION OF CONFORMITY

SILVERTRONIC Ltd. confirm that the products listed below have been Designed, Manufactured and Tested to, and therefore are fully compliant to; *IEC1010-1 1990 inc. Amendment 2-1995 & IEC1010-2-031 First Edition-1993* and thus meet THE LOW VOLTAGE DIRECTIVE (LVD) 73/23/EEC and amending directive 93/68/EEC

PRODUCT TYPE / DESCRIPTION				Test Lead Set	
PART No.				129526	
IEC1010 RATING	VOLTAGE	1000	Volts	INSULATION CATEGORY	III (3)
***126017	VOLTAGE	300	Volts	INSULATION CATEGORY	II (2)

MAXIMUM RATED CURRENT	**10/*20/30	Amps
POLLUTION DEGREE	2	
INSULATION CATEGORY	DOUBLE/REINFORCED	

INCLUDES FOLLOWING ASSEMBLIES / COMPONENTS SUBJECT TO INDIVIDUAL TEST.	DESCRIPTION	PART No.
	*OEM Full Safety Plug To OEM Right Angle Plug Lead	123815
	OEM Style Prod	125072
	1000V Croc. Clip	126024
	*Sprung Hook	121026
	**Thin Spikes	121039
	***Small Croc. Clip	126017



All Testing carried out by Third Party Testhouse CRANAGE EMC TESTING Ltd. (LVD TESTING DEPARTMENT)



www.silvertronic.com e-mail: sales@silvertronic.com

Registration No. 1993292

North American Office

4405 Sharon Copley Road, Medina, Ohio, 44256. Toll Free: (866) 556 3308 Facsimile: (330) 722 2281

