

## DECLARATION OF CONFORMITY

SILVERTRONIC Ltd. confirm that the products listed below have been Designed, Manufactured and Tested to, and therefore are fully compliant to;  
**IEC1010-1 1990 inc. Amendment 2-1995 & IEC1010-2-031 First Edition-1993**  
and thus meet **THE LOW VOLTAGE DIRECTIVE (LVD) 73/23/EEC and amending directive 93/68/EEC**

PRODUCT TYPE / DESCRIPTION				4mm FULL SAFETY SPADE TERMINALS		
PART No.				134206		
IEC1010 RATING	VOLTAGE	1000	Volts	INSULATION CATEGORY	III ( 3 )	

MAXIMUM RATED CURRENT	20	<b>Amps</b>
POLLUTION DEGREE	2	
INSULATION CATEGORY	DOUBLE/REINFORCED	

INCLUDES FOLLOWING ASSEMBLIES / COMPONENTS SUBJECT TO INDIVIDUAL TEST.	DESCRIPTION	PART No.
	4mm SPADE TERMINAL	124000



All Testing carried out by Third Party Testhouse  
**CRANAGE EMC TESTING Ltd.**  
( LVD TESTING DEPARTMENT )



www.silvertronic.com  
e-mail: sales@silvertronic.com

Registration No. 1993292

**North American Office**  
4405 Sharon Copley Road,  
Medina, Ohio, 44256.  
Toll Free: (866) 556 3308  
Facsimile: (330) 722 2281



Date  
22/12/98

Drn.By  
Steve S

Title FULL SAFETY SPADE

ISSUE No.

15/07/97

1

PROD. ISSUE

Part No.


134206

This drawing is not to be disclosed to a third party without the authority from the company.

Scale  
FULL SIZE


IEC 1010 COMPLIANT

I No. RED & BLACK 124000



MOULDING 113327

NEW CONTACT 103105/N



108000/N SPADE

ASSEMBLY JIG NUMBERS  
124000=J156.158

PACKAGING 111503 Poly Bag  
LABEL

NO PRINTING

TOLERANCES

CABLE LENGTHS +/- 5%

DO NOT SCALE.  
ALL DIMENSIONS IN mm.