
Laboratory DC Power Supply





Features:

- Maximum rating 30 V, 3 A.
- · Fine and coarse adjustments for accurate voltage and current settings.
- Constant voltage and constant current operations.
- HI / LO current range selector.
- Series or parallel operation.
- Low ripple and noise.
- Overload and reverse polarity protection.
- Power requirements : 100 V, 120 V, 220 V, 240 V ac, 50/60Hz (selectable).
- Includes test leads and owner's manual.
- One year limited warranty.
- Dimensions : 128 (W) × 145 (H) × 285 (D) mm.
- Weight : 5 kg.

[™]Warning :

Voltages more than 60 V dc are a lethal shock hazard to the user. Be Careful when connecting power supplies in series to achieve voltages higher than 60 V dc total or 60 V dc between any connection and earth ground.

www.element14.com www.farnell.com www.newark.com www.cpc.co.uk www.mcmelectronics.com



Laboratory DC Power Supply

: 0 to 30 V dc continuously adjustable

 $\therefore \le 0.01\% + 3 \text{ mV}$ (rating current $\le 3 \text{ A}$) $\therefore \le 0.01\% + 5 \text{ mV}$ (rating current > 3 A)

: 0 to 3 A continuously adjustable

: Line regulation $\therefore \le 0.2\%$ + 3 mA

 $1 \le 100 \ \mu s$ (50% load change, minimum load 0.5 A)

 $:\le 0.5$ mV rms (5 Hz to 1 MHz) (rating current \le 3 A) $:\le 1$ mV rms (5 Hz to 1 MHz) (rating current > 3 A)

: ≤ 0.01% + 3 mV

: ≤ 300 PPM/°C

: ≤ 0.2% + 3 mA

 $: \le 3 \text{ mA rms}$



Specifications :

Constant Voltage Operation:

Output voltage range

Voltage regulation:

Line regulation Load regulation Load regulation Recovery time Ripple and noise

Temperature coefficient

Constant Current Operation:

Output current range Current regulation Load regulation Ripple and noise

Display:

Analog - 1 each voltmeter and ammeter	
Class	: 2.5
Size	: 50 × 50 mm

Insulation:

Between chassis and output terminal 20 megaohms or above (500 V dc) Between chassis and AC cord 30 megaohms or above (500 V dc)

Environmental:

: 0°C to 40°C, < 80%
: 10°C to 70°C, < 70%

Part Number Table

Description	Part Number
Laboratory DC Power Supply	72-2010 CALIBRATED

Important Notice : This data sheet and its contents (the "Information") belong to the members of the Premier Famell group of companies (the "Group") or are licensed to it. No licence is granted for the use of it other than for information purposes in connection with the products to which it relates. No licence of any intellectual property rights is granted. The Information is subject to change without notice and replaces all data sheets previously supplied. The Information supplied is believed to be accurate but the Group assumes no responsibility for its accuracy or completeness, any error in or romasion from it or for any use made of it. Users of this data sheet sheuld check for themselves the Information or use of it (including liability resulting from negligence or where the Group was aware of the possibility of such loss or damage eristing) is excluded. This will not operate to limit or restrict the Group's liability for death or personal injury resulting from its negligence. Tenma is the registered trademark of the Group. © Premier Farnell plc 2011.

www.element14.com www.farnell.com www.newark.com www.cpc.co.uk www.mcmelectronics.com

