


FED-26852-106

SHEET 2.1

THE FED-26852-106 IS A FERROFLUID AND TYPE III DAMPED RECEIVER WITH EXTERNAL VENT AND A PEAK OF 6dB RELATIVE TO THE SENSITIVITY AT IKHZ UNDER CONSTANT VOLTAGE CONDITIONS. THS MODEL HAS INCREASED DCR / IMPEDANCE RATIO.





FREQUENCY IN HERTZ

## **ACOUSTICAL**

SENSITIVITY DEVICE WILL PRODUCE THE SPL LISTED BELOW UNDER TEST CONDITIONS DESCRIBED IN TABLE 3. NOMINAL SENSITIVITY AT IKHZ IS dB RELATIVE TO 20µPa. ALL OTHER VALUES IN dB RELATIVE TO THE SENSITIVITY AT IKHZ.

FREQUENCY (Hz)	MINIMUM	NOMINAL	MAXIMUM
200	-3.0	0.0	+3.0
500	-1.0	+1.0	+3.0
1000	-2.0	102.5	+2.0
2000 - 3000	+3.0	+6.0	+9.0
4000	-12.0		

PORT LOCATION: 12N

TABLE I

TOTAL HARMONIC DISTORTION DEVICE WILL NOT EXCEED TOTAL HARMONIC DISTORTION LEVELS LISTED BELOW.

FREQUENCY (Hz)	AC DRIVE (V rms)	DC BIAS (mA)	LIMIT (%)
500	1.02	0	10
830	0.362	0	5
1250	0.362	0	5

TABLE 2

## TEST CONDITIONS

NOMINAL SOURCE VOLTAGE	0.362 V rms, OmA DC BIAS
SOURCE IMPEDANCE	< I Ohm
TUBING	
COUPLER CAVITY	2 CM <sup>3</sup> , SIMULATED ANSI S3.7 TYPE HA-3 (IEC 126)
TABLE 3	

## **ELECTRICAL**

DC RESISTANCE	201 OHMS ± 10%
IMPEDANCE @ 500 Hz	280 OHMS ± 15%
IMPEDANCE @ I kHz	408 OHMS ± 15%

ISOLATION: CASE WILL BE ELECTRICALLY ISOLATED FROM THE COIL CIRCUIT.
BAROMETRIC RELIEF: THE AIR FLOW THROUGH THE BAROMETRIC FELIEF VENT TO BE 1.0 TO 2.5 CC/MIN WHEN AIR PRESSURE OF 15 INCHES WATER (3736 Pa) IS APPLIED TO THE PORT APERTURE.

TEMPERATURE: OPERATING RANGE FROM 0°C TO 63°C (SENSITVITY WILL NOT VARY BY MORE THAN ±3 dB WITHIN RANGE)
SENSITIVITY AT 0°C IS 2dB LOWER THAN THE SENSITIVITY AT ROOM TEMPERATURE.

DELIA PEAK IS	IdB HIGHER	AT BODY	TEMPERATURE	(3/°C)
CTODACE DANCE	EDOM - 40°C	TA 620 C		

Revision	C.O. #	Implementation Date	RELEASE LEVEL		REVISION
D	C10107990	11-13-08	Active		D
WHEN TEST LIMITS ARE USED TO ESTABLISH INCOMING INSPECTION ACCEPTANCE/REJECTION CRITERIA, CORRELATION OF TEST EQUIPMENT WITH KNOWLES IS ALSO REQUIRED FOR				DR. BY	DATE  - 3-06
ELIMINATIO	ELIMINATION OF EQUIPMENT AND TEST METHOD VARIATION			CK. BY	DATE
TITLE:	RE	CEIVER	FED-26852-106	GJP APP BY	1 - 16 - 06
	PERFORMAI	NCE SPECIFICATION	SHT 2.1	GJP	1-16-06

## KNOWLES ELECTRONICS ITASCA, ILLINOIS U.S.A.